

TENDER FOR
SUPPLY AND INSTALLATION OF MICRO COMPUTED TOMOGRAPHY
AT
NATIONAL INSTITUTE OF AYURVEDA, JAIPUR

NIT ISSUE DATE	:	07-01-2022 (3:00 PM)
NIT NO.	:	NIA/CS/AH/(1)87/2021
LAST DATE OF SUBMISSION	:	17-01-2022 (2:00 PM)
BID OPENING DATE	:	18-01-2022 (2:00 PM)
TENDER FEES	:	2500/-



NATIONAL INSTITUTE OF AYURVEDA
(Deemed to be University)
MINISTRY OF AYUSH GOVT. OF INDIA
JORAWAR SINGH GATE AMER ROAD, JAIPUR

Supply and Installation of Micro Computed Tomography
NIA/CS/AH/(1)87/2021

National Institute of Ayurveda, Jaipur an autonomous body under the Ministry of AYUSH, Government of India invites **online bids in two bid system of tender** for **Supply and Installation of Micro Computed Tomography**. Best offers are invited along with the complete details of specifications, terms & conditions.

S.N.	ITEM DESCRIPTION	QTY	EMD (RS.)
1	Supply and Installation of Micro Computed Tomography	01	4,00,000/-

INSTRUCTIONS:

1. Bids shall be submitted online at CPPP website: <https://eprocure.gov.in/eprocure/app>
2. The complete bidding process is online Bidders should be in possession of valid digital Signature Certificate (DSC) of class II or III for online submission of Bids, Prior to bidding DSC need to be registered on the website mentioned above.
3. Tenderer/Contractor/Bidders are advised to follow the instructions provided in the instructions to the contractors/Tenderer/Bidders for the e-submission of the bids online through the Central Public Procurement Portal for e-Procurement at <https://eprocure.gov.in/eprocure/app>.
4. Bid documents may be scanned with 100 dpi in black and white option which helps in reducing size of the scanned document.
5. Tenderer /Contractor/Bidders are advised to submit Technical Bid Document photo copy before last date of bid submission.
6. **EMD Payment:**

The Bidder shall be required to submit the Earnest Money Deposit (EMD) for an amount of **Rs. 4,00,000 (Rupees Four Lacs Only)** by the way of demand drafts only. The demand drafts shall be drawn in favor of **“Director, National Institute of Ayurveda”** Payable at **Jaipur**. The EMD of the Successful Bidder shall be returned after the successful submission of Bank Guarantee/Security Deposit and for unsuccessful Bidder(s) it would be returned after award of the Contract. **The demand Drafts for EMD must deliver to Director, NIA, Jaipur on or before last date /time of Bid Submission.**

 - a) Tenderer shall not be permitted to withdraw his offer or modify the terms and conditions thereof. In case the tenderer fails to observe and comply with stipulation made herein or back out after quoting the rates, the aforesaid amount of earnest money will be forfeited.
 - b) The firm who are registered with National Small industries corporation (NSIC) or Small Scale Industries (SSI) are exempted to submit the EMD(Copy of registration must be provided along with Technical Bid) Registration proof and Udyog Aadhar Certificate should be enclosed with Technical Bid.
 - c) Bidders are not allowed to submit more than one bid anytime during the tendering process for the same/similar tendered item else all his bids shall be cancelled thereby making him disqualified in addition to the forfeiture of the EMD.
 - d) The EMD, in case of unsuccessful Bidders shall be retained by NIA, Jaipur till the finalization of the Tender. No Interest will be payable by NIA, Jaipur on EMD.
 - e) EMD is required to protect the purchaser against the risk of Bidders conduct. The EMD will be forfeited if the bidder withdraws or amends its tender or impairs or derogates from the tender in any respect within the period of validity of its tender or if it comes to the notice that the information/ documents furnished in its tender is incorrect or false.

Supply and Installation of Micro Computed Tomography
NIA/CS/AH/(1)87/2021

- f) The EMD/PBG shall be forfeited if successful bidder fails to supply the goods/equipment in stipulated time or fails to comply with any of the terms & conditions of the contract or fail to sign the contract.

7. Submission of Tender:

The tender shall be submitted online in two parts, viz., Technical Bid and Financial Bid (BOQ). All the pages of bid being submitted must be signed and sequentially numbered by the bidder irrespective of nature of content of the documents before uploading.

The Bidder shall quote the Technical and Financial Bids as per the format enclosed with tender document. ***The Bidders should submit their all relevant documents (Technical bid document and EMD proof etc.) before last date and time of bid submission.***

The offers submitted by Telegram/Fax/email shall not be considered. No correspondence will be entertained in this matter.

8. Technical Bid:

All pages of the Tender should be numbered and indexed.

The bidder shall provide in its tender the required as well as the relevant documents like technical data, literature, drawings etc. to establish that the goods and services offered in the tender fully confirm to the goods and services specified by the purchaser in the tender documents. For this purpose, the bidder shall also provide a clause-by-clause commentary on the technical specifications and other technical details incorporated by the purchaser in the tender documents to establish technical responsiveness of the goods and services offered in its tender duly indicating relevant page numbers in the product literature.

The following documents are to be furnished by the contractor/Bidder along with Technical Bid as per the Tender Document:

- a) Signed and scanned copy of appropriate value of valid registration certificate (if any), experience certificate as per the tender notice, PAN, GST registration certificate and Tender Acceptance Letter.
- b) Signed and Scanned Copy of make and model of all systems, sub systems and additional items should be mentioned in the technical bid and complete technical details should be provided in the form of Brochures and write-ups.
- c) The bidders are required to submit user certificate for the relevant equipment on the letter head of the institution (Government/ Private)
- d) Signed & stamped compliance sheet of the technical specification of the goods with technical printed literature must be enclosed with the bid
- e) **Manufacturer Authorization:**
The bidder (if not original equipment manufacturer must submit Original Equipment Manufacturer authorization certificate that the tenderer is authorized for selling and maintain the equipment quoted for.

9. Financial Bid:

Price Schedule(s) as per BoQ format filled up with all the details including Make, Model etc. of the goods offered to be uploaded.

While filling up the columns of the Financial Bid, the following aspects should be noted for compliance:

Supply and Installation of Micro Computed Tomography
NIA/CS/AH/(1)87/2021

Price should include:

- a) Duties and fees, if any
- b) Charges towards Packing & Forwarding, Inland Transportation, Insurance, Loading/Unloading and other local costs incidental to delivery of the goods to their final destination as specified in the List of Requirements and Financial Bid
- c) The price of annual CMC, as mentioned in List of Requirements, Technical Specification and Financial Bid
- d) GST will be payable separately as per applicable rate.

Schedule of price bid in the form of BOQ_XXXX .xls

The below mentioned price bid format is provided as BoQ_XXXX.xls along with this Tender Enquiry Document at <https://eprocure.gov.in/eprocure/app> . Bidders are advised to download this BoQ_XXXX.xls as it is and quote their offer/rates in the permitted column and upload the same in the commercial bid. Bidder shall not tamper/modify downloaded price bid template in any manner. In case if the same is found to be tempered /modified in any manner, tender will be completely rejected and tenderer is liable to be banned from doing business with NIA, Jaipur.

10. The authorized signatory of the bidder must digitally sign the bid. Individuals digitally signing the bid or other documents connected with a contract must specify whether he signs as:
 - a) A 'Sole Proprietor' of the firm or constituted attorney of such Sole Proprietor
 - b) In case of partnership firm he must have authority to quote & to refer to arbitration dispute concerning the business of the partnership either by virtue of the partnership agreement or a power of attorney
 - c) Constituted attorney of the firm if it is a company.

11. Turnover

The Firm/Agency should have an annual minimum turnover of Rs. 2 Crore during last 3 Years (2018- 19, 2019-20, 2020-21) to be eligible for consideration.

12. Validity:

The quoted rates must be valid for a period for 180 days from the date of closing of the tender. The overall offer for the assignment and bidder(s) quoted price shall remain unchanged during the period of validity. If the bidder quoted the validity shorter than the required period, the same will be treated as unresponsive and it may be rejected.

13. The bidding firm should be continuously engaged (with Valid License/Registration) in the same business at least for last 3 Years. The bidding firm should have at least 1 similar contract with Central Government/State Government/Public Sector Undertakings/ Universities/ Deemed to be universities/ Autonomous Bodies/Reputed Medical Institute or Organizations in last 3 Years.
14. The bidder shall provide a list of major Government and Private Institutions where its relevant bid item has been supplied during last 3 years along with order copies
15. **The bidder firm should not have been blacklisted by any Ministry/Department of Govt. of India/State Government/any PSU's etc. The Performance Security Deposit will be forfeited after awarding the Bid, in case the bidding firms found black listed by Central Government/State Government/PSU's/Universities/Deemed to be Universities/Autonomous Bodies at any point of time.**
16. **The Director, National Institute of Ayurveda reserve the right to accept or reject any/all tenders without assigning any reason thereof.**

Supply and Installation of Micro Computed Tomography
NIA/CS/AH/(1)87/2021

17. Availability of Funds:

Expenditure to be incurred for the proposed purchase will be met from the funds available with the purchaser/consignee

18. Delivery and Installation:

For goods supplied from India:

- a) All the goods ordered shall be delivered and Installed at NIA, JAIPUR within **30 days** from the date of issue of supply order. All the aspects of safe delivery, installation and commissioning shall be the exclusive responsibility of the supplier. If the supplier fails to deliver, install and commission the goods on or before the stipulated date, then a penalty at the rate of 0.5% per week or a part thereof of the total order value shall be levied subject to maximum of 10% of the total order value. The successful tenderer will also provide required training for supplied items at NIA.JAIPUR. The goods should be manufactured after adoption of latest technology. If at any time during the currency of the contract, the supplier encounters conditions hindering timely supply of the goods and performance of services, the supplier shall promptly inform the NIA, JAIPUR for extension of the delivery schedule accordingly.
- b) On receiving the supplier's communication, NIA, JAIPUR shall examine the situation as soon as possible and, at its discretion, may agree to extend the delivery schedule, with or without liquidated damages for completion of supplier's contractual obligations by issuing an amendment to the contract. In the case of package supply where the delayed portion of supply materially hampers installation and commissioning of the systems, liquidated damages charges shall be levied as above on the total value of the concerned package of the purchase order. Quantum of liquidated damages assessed and levied by the purchaser shall be final and not challengeable by the supplier.

19. Signing the Contract:

The successful bidder shall be required to execute the Contract Agreement accepting all terms and conditions stipulated herein on a non-judicial stamp paper of Rs. 500/- (Rs. Five Hundred only) along with performance security within fifteen days of the issue of the Letter of notification of award. In the event of failure on the part of the successful bidder to sign the Contract within the period stipulated above, the EMD shall be forfeited and the acceptance of BID shall be considered as cancelled.

20. Performance Security:

As a guarantee towards due performance and compliance of the contract work, the successful bidder (contractor) will deposit an amount equal to 3% of order value and should be kept valid for a period of 60 day beyond completion of all the contractual obligation, including CMC period towards security deposit by way of demand draft/ bank Guarantee in favour of "**DIRECTOR NIA, JAIPUR**" payable at JAIPUR drawn on any Nationalized Bank/Scheduled Bank and payable at JAIPUR within fifteen days of the issue of the Letter of notification of award along with non-judicial stamp paper of Rs. 500/- (Contract agreement).

21. Incidental Services:

The supplier shall be required to perform the following services: -

- a) Installation & Commissioning, Supervision and Demonstration of the goods.
- b) Providing required jigs and tools for assembly, minor civil works required for the completion of the installation.
- c) On Site Training to Doctors/ Technicians/ Staff is to be provided by Supplier for operation and maintenance of the equipment for a period of 30 working days after successful installation of the machine, as per direction of user department.
- d) Supplying required number of operation & maintenance manual for the goods.

Supply and Installation of Micro Computed Tomography
NIA/CS/AH/(1)87/2021

- e) To provide non-locked open software and standard interface inter-operability conditions for networked equipment's in hospital management information system, wherever applicable.

22. After Sales Service:

After sales service Centre should be available on 24 (hrs.) X 7 (days) X 365(days) basis. Complaints should be attended properly, maximum within 24 hours to ensure an uptime of minimum 95%, wherever applicable, failing which the necessary penalty measures shall be enforced.

23. Inspections:

- a) NIA, Jaipur shall have the right to inspect and/or to test the goods to confirm their conformity to the NIT Specifications at no extra cost to the Purchaser.
- b) NIA, Jaipur right to inspect, test and, where necessary, reject the Goods after the goods arrival at the final destination shall in no way be limited or waived by reason of the Goods having previously been inspected, tested and passed by NIA,JAIPUR prior to the goods shipment.
- c) The Director, NIA,JAIPUR shall be the final authority to reject full or any part of the supply which is not confirming to the specification and other terms and conditions
- d) No payment shall be made for rejected items and those must be removed by the Bidders within two weeks of the date of rejection at their own cost and replaced immediately. In case these are not removed, these will be auctioned at the risk and responsibility of the suppliers without any further notice.

24. Payment Terms:

Payment of the total order value shall be released after the successful installation/ commissioning and tanning of the ordered goods against the submission of the inspection report.

25. Guarantee / Warrantee Period:

The Tenderers must quote for 5 year comprehensive warranty (including all Spares, Accessories and Labor) from the date of completion for satisfactory installation. The Warranty Charges shall not be quoted separately otherwise the offer shall be summarily rejected.

26. Breach of Terms and Conditions:

In case of breach of any terms and conditions as mentioned above, the Competent Authority, will have the right to cancel the work order/ job without assigning any reason thereof and nothing will be payable by NIA, Jaipur in that event the security deposit shall also stands forfeited

27. Insolvency etc.:

In the event of the firm being adjudged insolvent or having a receiver appointed for it by a court or any other order under the Insolvency Act made against them or in the case of a company the passing any resolution or making of any order for winding up, whether voluntary or otherwise, or in the event of the firm failing to comply with any of the conditions herein specified NIA, Jaipur shall have the power to terminate the contract without any prior notice

28. Force Majeure:

If, at any time during the subsistence of this contract, the performance in whole or in part by either party of any obligation under this contract is prevented or delayed by reasons of any war or hostility, act of public enemy, civil commotion, sabotage, fire, floods, explosion, epidemics,

Supply and Installation of Micro Computed Tomography
NIA/CS/AH/(1)87/2021

quarantine restriction, strikers lockout or act of God (herein after referred to as events) provided notice of happening of any such eventuality is given by part to other within 21 days from the date of occurrence thereof, neither party shall by reason of such event be entitled to terminate this contract nor shall either party have any claim for damages against other in respect of such non-performance or delay in performance, and deliveries have been so resumed or not shall be final and conclusive. Further, that if the performance in whole or in part of any obligation under this contract is prevented or delayed by reason of any such event for a period exceeding 60 days, either party may, at least option to terminate the contract.

29. Bidder shall submit a copy of the tender document and addenda thereto, if any, with each page of this document should be signed and stamped to confirm the acceptance of the entire terms & conditions as mentioned in the tender enquiry document.

30. After due evaluation of the bid(s) Institute will award the contract to the lowest evaluated responsive tenderer.

31. Conditional bid will be treated as unresponsive and it may be rejected.

32. Demonstration:

NIA, Jaipur reserves the right to ask the tenderers for arranging demonstration of their equipment for which rates have been quoted, to the concerned committee, if required.

33. The Institute reserves the right to accept in part or in full or reject any or more tender(s) without assigning any reasons or cancel the tendering process and reject all tender(s) at any time prior to award of contract, without incurring any liability, whatsoever to the affected bidder or bidder(s)

34. Applicable Law:

The contract shall be governed by the laws and procedures established by Govt. of India, within the framework of applicable legislation and enactment made from time to time concerning such Commercial dealings / processing

Any disputes are subject to exclusive jurisdiction of Competent Court and Forum in Jaipur, Rajasthan, India only

The Arbitration shall be held in accordance with the provisions of the Arbitration and Conciliation Act, 1996 and the venue of arbitration shall be at Jaipur. The decision of the Arbitrator shall be final and binding on both the parties

Force Majeure: Any delay due to Force Majeure will not be attributable to the supplier.

Supply and Installation of Micro Computed Tomography
NIA/CS/AH/(1)87/2021
Technical Bid

1.	Name of Tendering Company/Firm/Agency	:	
2.	Name of Owner/Partners/Directors	:	
3	Full Particulars of the Head Office	:	
	• Address	:	
	• Telephone No.	:	
	• Fax No.	:	
	• E-mail address	:	
4	Full Particulars of the Bankers of the Company/Firm/Agency, with full Address/Tel. No.	:	
	• Name of the Bank	:	
	• Address of the Bank	:	
	• Telephone No.	:	
	• Fax No.	:	
	• E-mail address	:	
5	Registration Details:	:	
	• Copy of Firm/Agency's Registration No.	:	
	• PAN/GIR NO.	:	
	• GST Registration No	:	
	• Aadhar No. of Owner/Partners/Directors	:	
6	Details of Earnest Money Deposit	:	
	• Amounts (Rs.)	:	
	• DD/PO No. and Date	:	
	• Drawn on Bank	:	
	• Valid upto	:	
7	Details of Tender Fee	:	
	• Amount	:	
	• DD/PO No.	:	
	• Drawn on Bank	:	
	• Valid upto	:	
8	Annual Turnover of the Company/Firm/Agency for the last 3 years (With Proof duly certified by Chartered Accountant/Auditor) Last 03 Years balance sheet and copies of Income tax returns for 3 Years as stated in Terms & Conditions		
	2018-2019	:	
	2019-2020	:	
	2020-2021	:	
9	The Firm who are registered with NSIC or Small Scale Industry/Or MSME are exempted to submit the tender fee or EMD (Copy of Registration certificate must be provided along with Udyog Addhar)		
10	Similar supplies in last 3 Years (with Proof)		
	Experience in the field for the minimum 3years		1. Similar 1 Government Running Contract with Central Government /State Government/Public Sector Undertakings /Universities/ deemed to be universities / Autonomous Bodies / Reputed Medical Institute or Organizations in last 3 Years.

Supply and Installation of Micro Computed Tomography
NIA/CS/AH/(1)87/2021

Certificate
(To be submitted with the Technical Bid)

The Annual Turnover of M/s. _____ for the last 3 Years are given below and it is certified that the Statement is true and correct:

S.N.	Year	Turnover (in Rs.)
1	2018-2019	
2	2019-2020	
3	2020-2021	
	Total	Rs _____ Lacs

Average Turnover per Annum: Rs. _____ Lacs.

Signature of Chartered Accountant/Auditor with Seal

Supply and Installation of Micro Computed Tomography
NIA/CS/AH/(1)87/2021

MANUFACTURER'S/PRINCIPAL'S AUTHORIZATION FORM

TO,
Director
National Institute of Ayurveda
Jaipur
Sir,
Tender:-----.

We ----- who are established and reputable manufacturer of having factories at ----- and -----, hereby authorize Messrs.----- (name and address of agent(s) to bid, negotiate and conclude the contract with you against Tender.

No----- for the above goods manufactured by us. No company or firm or individual other than Messrs. ----- are authorized to bid, negotiate and conclude the contract in regards to this business against this specific tender.

We hereby extend our full guarantee and warranty as per the conditions of tender for the goods offered for supply against this tender by the above firm.

The authorization is valid upto -----

Yours faithfully,

For and on behalf of Messrs.'-----
(Name of manufacturers) principal

Supply and Installation of Micro Computed Tomography
NIA/CS/AH/(1)87/2021

Technical Specifications:

Specification for Micro- CT (In-vivo)

Qty: 01 Pc

S. NO.	Parameter	Specification
1	X-ray source	Continuous variable peak energy from 40 till 100kV, < 5 micron spot size at 4 watt; 20W maximum power, air cooled sealed type.
2	X-ray detector	Largest field of view cooled 16Mp (4096 x 4096) sCMOS detector
3	Filter changer	Al 0.5mm, Al 1mm, Al 0.5mm + Cu 0.03mm, Cu 0.1mm and "low-dose" filter (beam shaper).
4	Low contrast spatial resolution	<6 µm (10%MTF)
5	Scanning diameter	12 to 80mm dependent on magnification.
6	Scanning length	More than 300mm by multiple, automatically connected scans, 25mm in one scan or >250mm by single spiral (helical) scan.
7	Reconstruction arrays	Cross section format (isotropic grid): Up to 4000 x 4000 for up to 40mm scanning diameter and up to 8000 x 8000 pixels for up to 80mm scanning diameter. The scanner includes Physiological Monitoring subsystem for real-time visual monitoring, movement detection, synchronization with breathing and heart activities and keeping stable preselected animal temperature.
8	Radiation safety	< 1 microSv/h average during full scan in 10cm from the instrument surface.
9	Animal Cassettes	Number: Three, 28mm in diameter for small mice, 40mm in diameter for large mice and small rats and 70mm in diameter for large rats, with connections to ECG electrodes, temperature sensor and anaesthetic gas connected to face mask.
10	Sample holders	Number: Two, 20 mm in diameter, 150 mm in length and 12 mm in diameter, 120mm in length. Large carbon composite bed, 73mm internal diameter / 76mm external diameter, 300mm in length Foam holders for ex-vivo and peripheral scanning. Non-metallic (carbon fibre based) ECG electrodes and temperature sensor.
11	Software	Good laboratory Practices compliance Software with feature of Source and camera control, obtaining shadow images and series of angular projections for reconstruction, acquisition parameter control including different sets of measurement protocols. Full cross section size (full image mode), partial reconstruction mode, and possibility for detail local reconstruction with object larger than field of view. Speedups of 10x to >100x (higher for large image formats) compare to best other reconstruction software and 2x to 10x compare to GPU-accelerated reconstruction. Images are displayed as a slice-by-slice movie. DataViewer also has a viewing mode with three intersecting orthogonal slices centered at any selected point inside the reconstructed volume, which can be turned and each intersecting slice independently moved by simple mouse control. Measures comprehensive 2D and 3D morphometric parameters. The

Supply and Installation of Micro Computed Tomography
NIA/CS/AH/(1)87/2021

		<p>software features an advanced, rapid and user-friendly region-of-interest (ROI) free-hand selection tool, which by autointerpolation between layers becomes the volume-of-interest (VOI), the essential basis for all quantitative analysis. CT density analysis is also provided within any VOI, which by calibration can provide Hounsfield units or volumetric bone mineral density. The 3D models created by CT-analyzer are loaded and viewed from software.</p> <p>Realistic surface rendering visualization software CT-volume provides viewing, intuitive manipulation and flexible control of 3D-model texture and transparency, plus a versatile moviemaker. CT-volume loads the triangulated models created in CT analyzer, in different formats including STL.</p>
12	Phantoms for bone research	<p>One pair of 2 mm diameter rods (250 & 750 mg/cm³) for mouse bone calibration and one pair of 4 mm diameter rods (250 & 750 mg/cm³) for rat bone calibration.</p> <p>Aluminium foils for trabecular thickness calibration. The phantom consists of a stack of 4 foils with a thickness of 20, 50, 125 and 250 micron. The foils are separated by plastic film.</p> <p>Standard set of 4mm/2mm phantoms + Al foils phantom (1 set), 8-16mm phantoms (1 set), 32mm phantoms (1 set), Expanded poly-styrene tube set. 6 Types each 10 nos (1 set), Poly-propylene foam tube set. 3 types each 10 nos. (1 set)</p>
13	Computer	<p>Dual Intel XEON processor E5-2640 v4 (ten core HT, 3.4 GHz Turbo 25 MB, 90 W), Windows 10 Professional English (64-bit) (Windows 7 if required / until available), 128 GB DDR4 2400 MHz RAM, NVIDIA® Quadro® Graphical card, 12 TB (3x4 TB) SATA HDD in RAID 0 for data, 512GB SolidStateDrive for operating system and programs DVD+/- RW drive, One 24" UltraSharp LCD/LED monitor with 1920 x 1200 pixels native resolution, Fire-Wire IEEE1394 card, QWERTY keyboard Optical, scroll mouse, Soundbar</p>
14.	Warranty	Comprehensive warranty for total 5 years on complete system including PM Kits and all hardware consumables is a must on complete system.